

Cleaning Instructions for HFA Inhalers

(Proventil, Ventolin, Proair, Xopenex)

- Remove medication canister!
- **Never get the canister wet.**
- Open the mouthpiece cap.
- Clean the plastic mouthpiece by running warm water through the top to the bottom for 30 seconds at least once a week.
- Shake to remove excess water, and then air-dry thoroughly (such as overnight).