

Cleaning Instructions for HFA Inhalers

(Flovent, Advair, Symbicort,)

- DO NOT REMOVE CANISTER!
- **Never get the canister wet.**
- Remove the mouthpiece cap.
- Use a cotton swab dampened with water to clean the small opening where the medicine sprays out of the canister.
- Repeat with a new dampened cotton swab.
- Wipe the mouthpiece with a damp cloth.
- Shake to remove excess water and air dry.
- Cleaning should be done once a week.